

**QUANTITATIVE TECHNICAL BID - Lot 1**

Open procedure for “Supply and assistance in the installation of two oceanographic buoys for the automatic measurement of sea meteorological parameters with control, management and remote transmission system of data”

**CIG 635699670C**

**CUP C63J13001100005**

The undersigned: \_\_\_\_\_ born on \_\_\_\_\_  
 in \_\_\_\_\_ (Province of \_\_\_\_\_), resident in No. \_\_\_\_\_,  
 Via \_\_\_\_\_, City \_\_\_\_\_  
 (Province of \_\_\_\_\_), in his capacity of Owner or Legal Representative of  
 \_\_\_\_\_ Company, with legal office in No. \_\_\_\_\_,  
 Via \_\_\_\_\_ City \_\_\_\_\_ (Province of \_\_\_\_\_)

being perfectly aware of all the prescriptions, official current rules and laws that govern the tender in question and all the clauses contained in the Call to Tender, in the Regulations, in the Special Specifications and, in particular, the mandatory prescriptions indicated in article 3 of the Special Tender Specifications.

**S T A T E S**

that, with reference to article 3 of the Special Specifications, the Technical Bid for LOT 1 of the above mentioned competitor Company is the following: *(cross out/describe for each assessment element, included between P1 and P12, the recurring hypothesis):*

	Assessment Elements	Technical Bid
<b>P1</b>	<u>Conductivity Sensor of which in Table 1: 1 more point will be assigned for an accuracy value lower than ±0.0005 (greater accuracy) (for a maximum total of 1 point)</u>	<input type="checkbox"/> Absence of the described characteristics <input type="checkbox"/> Presence of the described characteristics
<b>P2</b>	<u>Temperature Sensor of which in Table 1: 1 more point will be assigned for an accuracy value lower than ±0.05 (for a maximum total of 1 point)</u>	<input type="checkbox"/> Absence of the described characteristics <input type="checkbox"/> Presence of the described characteristics
<b>P3</b>	<u>Provision for additional sensors from Table 2, besides the 2 planned in point 16 of the Specifications: 0.5 more points will be assigned for every additional provision (for a maximum total of 3 points)</u>	Number of provision of additional sensors: _____

<b>P4</b>	<p><u>Extension of the period guarantee</u> beyond the planned 12 months: 5 more points will be assigned for 12 additional months (for a maximum total of 5 points)</p>	<input type="checkbox"/> Absence of the described characteristics <input type="checkbox"/> Presence of the described characteristics
<b>P5</b>	<p>SPECTRAL FLUOROMETER with measurement intervals from 1 to 200 microg chlorophyll a/l, 0.01 microg chla/l resolution, including cables , flow cell, with temperature and turbidity correction for the measurement of the chlorophyll and of the different algae classes: 15 more points will be assigned for the integration of the buoy of this tool (for a maximum total of 15 points)</p>	<input type="checkbox"/> Absence of the described characteristics <input type="checkbox"/> Presence of the described characteristics
<b>P6</b>	<p>INTEGRATION TO THE METEOROLOGICAL SENSORS with solar irradiance sensor (measurement range 0-1,250 W/m<sup>2</sup>) and UV-A and UV-B sensors (measurement range from 2 to 20 mW/cm<sup>2</sup>): 5 more points will be assigned for the supply of these sensors (for a maximum total of 5 points);</p>	<input type="checkbox"/> Absence of the described characteristics <input type="checkbox"/> Presence of the described characteristics
<b>P7</b>	<p>pH SENSOR with measurement interval 2-12 pH, 0.01 accuracy: 5 more points will be assigned for the supply of this sensor (for a maximum total of 5 points);</p>	<input type="checkbox"/> Absence of the described characteristics <input type="checkbox"/> Presence of the described characteristics
<b>P8</b>	<p>UNDERWATER TV CAMERA with the same characteristics of which in point 7) and related provision: 5 more points will be assigned for the supply of this TV camera (for a maximum total of 5 points);</p>	<input type="checkbox"/> Absence of the described characteristics <input type="checkbox"/> Presence of the described characteristics
<b>P9</b>	<p>UNDERWATER TV CAMERA with the same characteristics of which in point 8) and related provision: 5 more points will be assigned for the supply of these sensors (for a maximum total of 5 points);</p>	<input type="checkbox"/> Absence of the described characteristics <input type="checkbox"/> Presence of the described characteristics
<b>P10</b>	<p>DISSOLVED OXYGEN SENSOR with measurement interval 0-10 ml/l, 0.1 ml/l resolution: 5 more points will be assigned for the supply of this sensor (for a maximum total of 5 points);</p>	<input type="checkbox"/> Absence of the described characteristics <input type="checkbox"/> Presence of the described characteristics

<p><b>P11</b></p>	<p>Single-point acoustic CURRENT METER for current measurements with speed interval measurement from 0 to 300 m/sec, 0.1 mm/sec resolution, ±0.15 cm/sec accuracy; interval measurement for 0-360° direction, 0.01° resolution, ±8° accuracy, including wave-meter with interval measurement from 0 to 400 kPa (58 psia), &lt;0.0001 % FSO resolution, ±0.02 % FSO accuracy: 15 more points will be assigned for the integration in the buoy of this tool (for a maximum total of 15 points)</p>	<p><input type="checkbox"/> Absence of the described characteristics</p> <p><input type="checkbox"/> Presence of the described characteristics</p>
<p><b>P12</b></p>	<p>SENSOR FOR THE DETECTION OF HYDROCARBONS.: 5 more points will be assigned for the supply of this sensor (for a maximum total of 5 points);</p>	<p><input type="checkbox"/> Absence of the described characteristics</p> <p><input type="checkbox"/> Presence of the described characteristics</p>

Date, \_\_\_\_\_

\_\_\_\_\_

(The Owner of the Company or Legal Representative)

**Attach, under penalty of exclusion, photocopy of a suitable identification document of the signatory or, if applicable, of the signatories.**